



CERTIFICATE OF ANALYSIS

SILLIKER, Inc.
Southern California Laboratory
6360 Gateway Drive, Cypress, CA 90630
Tel. 877-777-6375 Fax. 714-226-0009
Email: getresults-sw@mxns.com

COA No:	SCA-43681509-0
Supersedes:	None
COA Date	6/25/20
Page 1 of 2	

Received From:	Watsonville, CA
Received Date:	6/19/20

Location of Test: (except where noted)
Cypress, CA

TO:
Mr. Zane Christopher
Lost Empire Herbs
195 Aviation Way
Watsonville, CA 95076

Analytical Results

Desc. 1:	GY080620	Laboratory ID:	400629922
Desc. 2:	GYNOSTEMMA	Condition Rec'd:	NORMAL
		Temp Rec'd (°C):	20.0
Analyte		Method Reference	Test Date Loc.
* - GMO Basic PCR Screening		BIOTECON PCR Kit Ins	6/25/20 CHG
P-35S	Not Detected	-	
P-FMV	Not Detected	-	
T-NOS	Not Detected	-	
LOD	0.01	%	
Sample Extracted	0.2	g	
Comment (LOD)	The limit of detection (LOD) of the method has been determined using certified reference material.		

Desc. 1:	SB090620	Laboratory ID:	400629928
Desc. 2:	SEABUCK	Condition Rec'd:	NORMAL
		Temp Rec'd (°C):	20.0
Analyte		Method Reference	Test Date Loc.
* - GMO Basic PCR Screening		BIOTECON PCR Kit Ins	6/25/20 CHG
P-35S	Not Detectable	-	
P-FMV	Not Detectable	-	
T-NOS	Not Detectable	-	
LOD	0.01	%	
Sample Extracted	0.2	g	
Comment (LOD)	The limit of detection (LOD) of the method has been determined using certified reference material.		
Comment (Plant)	Amplification control data indicate that plant DNA cannot be amplified because it is absent, damaged, or insufficient.		

Results reported herein are provided "as is" and are based solely upon samples as provided by client. This report may not be distributed or reproduced except in full. Client shall not at any time misrepresent the content of this report. Mérieux NutriSciences assumes no responsibility, and client hereby waives all claims against Mérieux NutriSciences, for interpretation of such results.

Except as otherwise stated, Mérieux NutriSciences Terms and Conditions for Testing Services apply.